

Department of BSBE

Corrigendum

To,
The Stores and Purchase Department,
IIT Kanpur

Sub: **Corrigendum** – Tender document, Technical specifications for Purchase of small volume spectrophotometer


Ref: Tender ref no. **IITK/BSBE/NM/2020/NC/09**

Dear Sir,

Please note that on page no.10, under the heading Technical Specifications, point no. 12 it is mentioned, Maximum Concentration: < 30,000 ng/μL (DNA).

Please change it to **Maximum Concentration: > 25,000 ng/μL (DNA)**

Sincerely,



Nitin Mohan

Dr. Nitin Mohan
Assistant Professor
Dept. of B.S.B.E.
I.I.T. Kanpur
Kanpur - 208016 (IND)
Tel: +91-0512-2594103