

**Indian Institute of Technology - Kanpur**  
**Department of Biological Sciences & Bioengineering**

Enquiry Number: **AB/BSBE/2017-2018/02**, dated: 16.05.2017

**Sub.: Inquiry for the Analytical Weighing Balance**

Opening date: 17 May, 2017 at 10:00 AM

Closing date: 29 May, 2017 at 5:00 PM

Sealed quotes (technical bid and price bid separately sealed) are invited for the **Analytical Weighing Balance** as per the specifications given in the next page.

**\*Your quote should mention/include the following:**

- Maximum discount if any should be offered and mentioned.
- Quoted price should include the cost for installation, warranty and required accessories (see below).
- Validity of the quote at least for 90 days.
- FOB (indicating port of shipment) and CIF (New Delhi) values should be quoted separately if import is required. For quotes in INR, the price quote should be for delivery at Kanpur.
- The quote should cover insurance for transport up to Kanpur.
- Indian agency commission if applicable (should be certified by the principal if no agency commission is applicable) in case of import.
- Authorization certificate from the principal if you are a local agent.
- Terms and conditions for the payment, including the banker's name of the principal and the account number, if any, for electronic transfer.
- Include proprietary item certificate if applicable.
- Technical literature to support your product (in technical bid).
- Users' list with contact address in technical bid.

The quote should reach the undersigned on or before 5 pm on or before 29 May, 2017.

The envelope should be marked as "**Analytical Weighing Balance**"

To,  
Dr Amitabha Bandyopadhyay  
Lab # 10,  
Department of Biological Sciences & Bioengineering  
Indian Institute of Technology,  
Kanpur 208016 (UP)

For any query, contact: Dr Amitabha Bandyopadhyay  
Department of Biological Sciences & Bioengineering  
Indian Institute of Technology,  
Kanpur 208016 (UP)  
Ph: 0512-259-4072

## “Analytical Weighing Balance”

**Essential components and features:**

**\*Item No. 1: Analytical Weighing Balance - (Quantity 1)**

**\*Specification:**

1	Maximum capacity	60-120 g
2	Minimum readability	0.01 mg
3	Minimum linearity	0.01 mg
4	Repeatability	≤ 0.08 mg
5	Settling time	Approx. 8 sec.
6	Calibration	Fully automatic time and temperature controlled internal adjustment
7	Regulatory requirements	ISO or any relevant trademark
8	Operation	Keys for main basic functions
9	Adjustment	Internal, isoCal
10	Overload protection	Must have

**\*Warranty period (from date of installation):**

Vendors providing 5years comprehensive warranty will be preferred