

# Request for quotation

Enquiry number: AE/KP/2016-17/Dehumidifier

Enquiry date: 21/10/2016

Closing Date: 31/10/2016

## **Subject: *Purchase of Dehumidifier (Make-NOVITA Singapore)***

Quotation for the item mentioned above is requested in a sealed envelope. The quotation should reach on or before October 31, 2016 to the address given below.

### **Technical specifications:**

- Power: 500-600 Watts
- Extraction: 30-35 liters/day will be sufficient
- Temperature: Should work in the range of 10-35 deg C
- Empty Weight: It should not be more than 15kg for easy handling
- Dimensions: (W x D x H): approximately 15in x 10in x 25in
- Tank Capacity: 6-8 liters capacity is sufficient
- Compressor: rotary preferred
- Coolant: Eco-friendly coolant
- Airflow: ~200-250 m<sup>3</sup>/hr
- Coverage area: ~3500 cu. ft.
- Noise Level: should be 50dB or lesser
- Further Preferred Feature:
  - a) Preferred humidity settings
  - b) Auto-stop with tank indicator
  - c) Auto humidity sensor

### **Terms and conditions:**

- Quotations should have a validity of minimum of 60 days
- The equipment should be provided with a warranty of at least one year after installation
- The delivery period should be specifically stated.
- Delivery period should not be more than 4 weeks and delivery should be at IIT Kanpur.
- Payments terms: 90% on installation and 10% satisfactory report
- Price must include all taxes and charges. All prices are to be FOR IIT Kanpur.

### **Address for the quotation:**

Dr. Kamal Poddar  
Department of Aerospace Engineering  
Indian Institute of Technology  
Kanpur-208016, India  
Phone: +91-512-2597843