

Tender No.: AS/CHE/2013-14/Aug/01

**Last Date: 27.08.13**

Subject : Quotation for Laboratory Vacuum Oven

Sealed quotation (separate technical and commercial bid) by post/courier is invited for the supply of Laboratory Vacuum Oven (1 Unit) as per the following specifications along with your complete terms and conditions. Your quote should be valid for minimum 90 days. Party must attach all the related certificates like Agency certificate, Proprietary item certificate (if applicable). Any quotation that does not carry technical and commercial bid separately or not accompanied with proper certificate will not be considered.

Specifications:-

- Double walled construction, inner chamber of size 45cm x 45cm x 45cm.
- Working chamber made of 3 mm thick Stainless steel plate of S.S.-304 grade.
- Corrosion resistant
- Four Nos of S.S.-304 Trays
- Chamber vacuum seal for full vacuum of 760mm of Hg.
- Temperature range from 50°C to 300°C.
- The Temperature accuracy in the range  $\pm 1\%$ .
- The Control Panel to consist of Mains MCB Switch, PID Controller, Vacuum Gauge, Vacuum Control Valve, and Exhaust Valve for purging inert gas, indicator lights, connecting cord and plug top.

Quotation must reach at the following address on or before August 27, 2013.

Prof. Ashutosh Sharma  
Department of Chemical Engineering  
Indian Institute of Technology Kanpur  
Kanpur 208016 INDIA