

INDIAN INSTITUTE OF TECHNOLOGY KANPUR
Department of 4i-Laboratory

Dr.Nachiketa Tiwari
Head
Department of 4i-Laboratory
IIT Kanpur
Kanpur 208016, U.P

Tel No:+91-512-6526/7475

Enquiry no.: 4i-Lab/2011-12/02

Enquiry Date: 03/02/2012
Closing date: 13/02/2012

Sealed quotations should reach the undersigned latest by 12.00 pm on 13/02/2012 for a fume-hood as described below.

Description

- Fume hood with low-noise, centrifugal blower (500 CFM) ducting 15 feet
- Chamber area: 5 feet wide, 30 inches deep and 8 feet high
- Working area: 4 feet width 30 inch depth & 30 inch height.
- Appropriate lighting inside the hood
- Three 15 A electrical sockets,
- Corrosion resistant construction
- Work surface to be acid and alkali proof
- Drawers below the work area to house equipment, etc.

Service features

Warranty period should be clearly mentioned in quotation. Local maintenance support should available. If not, please mention where the systems will be serviced in the event of failure.

You are requested to submit your quotation in sealed envelope with complete product description, technical literature, price, warranty period, delivery time and other conditions by February 13th 2012. Please note that your quotation must provide responses to all points raised under service features.



Dr.Nachiketa Tiwari

Head, 4i-Laboratory

IIT Kanpur-208016