

Civil Engineering Department, IIT Kanpur

DATED: 04/10/2012

Ref. No.: IITK/CE/EEM/MS/2012-13/Rajesh 1

Sealed or e-mailed quotations are invited by the undersigned with the following specifications for "Transparent FLUORINATED ETHYLENE PROPYLENE Hollow pipe roll": Last date for submission of quotation is upto 11/10/2012.

1. Material - FLUORINATED ETHYLENE PROPYLENE
2. Continuous Hollow pipe roll of above Material
3. Double layer width :31.5cm
4. Double layer thickness:0.16mm (80 micrometer for single layer).
5. Length: 125meters.
6. Transmittance should be greater than 90% for UV-a,b,c.
7. Material should be flexible and ductile.

Note:

1. All costs like material, transportation, all applicable taxes, etc. should be clearly stated.
2. Hard Copy or Scanned copy of Certificate of material, Photo of Material, quotation should be send to address as below.



Address:

Dr. Mukesh Sharma (Professor)

Civil Engineering Department

IIT Kanpur, Kanpur,

UP-208016, India

e-mail: mukesh@iitk.ac.in

Phone No.: +91-512-259-7759